

Hjertehjælp ApS
Rosenlund Sidevej 2
2791 Dragør
DENMARK

Eurofins Product Testing A/S
Smedeskovvej 38
8464 Galten
Denmark

CustomerSupport@eurofins.dk
www.eurofins.com

TEST REPORT

8 January 2021

1 Sample Information

Sample name	Test / Risen Medical / Batch no. 20201028
Sample reception	08/12/2020
Sample no.	392-2020-00561201
Analysis period	11/12/2020 - 08/01/2021

2 Picture of Sample



3 Results

Please see enclosure with detailed results.

4 Conclusion

The sample **complies** with the requirements in EN 14683 for a Type IIR surgical mask.

Eurofins Product Testing A/S



Pernille Krintel
Analytical Service Manager



Jeanette K. Pedersen
Analytical Service Manager

The results are only valid for the tested sample(s).

This report may only be copied or reprinted in its entirety, parts of it only with a written acceptance by Eurofins.

392-2020-00561201_FP_EN

5 Brief Evaluation of the Results

Type of analysis	Result	Standard
Bacterial Filtration Efficiency (BFE), (%)	99.81	EN 14683:2019, Annex B
Differential Pressure, (Pa/cm ²)	47.1	EN 14683:2019 + AC:2019, Annex C
Splash Resistance Pressure, (kPa)	> 16	ISO 22609:2004
Microbial Cleanliness (cfu/g)	< 30	EN 14683 / EN ISO 11737-1

Operational requirements for surgical masks based on EN 14683:2019 + AC:2019

Test	TYPE I	TYPE II	TYPE IIR
Bacterial Filtration Efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98
Differential Pressure, (Pa/cm ²)	< 40	< 40	< 60
Splash Resistance Pressure, (kPa)	Not required	Not required	≥ 16
Microbial Cleanliness (cfu/g)	≤ 30	≤ 30	≤ 30

*: Not accredited

<: Less than

>: Greater than

LOD: Limit of detection

Um(%): The expanded uncertainty Um(%) equals 2 x RSD%. For further information please visit www.eurofins.dk/uncertainty

The results are only valid for the tested sample(s).

This report may only be copied or reprinted in its entity, parts of it only with a written acceptance by Eurofins.

⌘: Internal test method

n.d: Not detected

n.m: Not measurable

LOQ: Limit of quantification

Analytical Report Nr.

AR-21-YL-000131-01

Sample code Nr.

560-2020-00009768

Date

07/01/2021

ANALYTICAL REPORT**Client Information**

Eurofins Product Testing Denmark A/S
Smedeskovvej 38
Galten DENMARK DENMARK

dk_usr_results_pt@eurofins.dk

For the attention of Interco results

Sample Information

Order Code: EUAA70-00009731
Reception Date: 14-Dec-2020
Analysis Starting Date: 14-Dec-2020
Analysis Ending Date: 7-Jan-2021
Sample code Nr. 560-2020-00009768
Sample described as: Masks

Requirements and decision rule

Customer requirements: EN 14683:2019+AC:2019 TYPE IIR
Decision Rule: Shared risk - Simple acceptance.

Information provided by the customer*

Client Reference: 392-2020-00561201
Sample Description: Ansigtsmasker Type IIR - True Nordic health
Purchase Order Number:

Batch 20201028

Analytical Report Nr.

AR-21-YL-000131-01

Sample code Nr.

560-2020-00009768

Date

07/01/2021

SAMPLE PICTURE

Analytical Report Nr.

AR-21-YL-000131-01

Sample code Nr.

560-2020-00009768

Date

07/01/2021

CONCLUSION:

TEST PROPERTY	PASS	FAIL	REMARKS
• Bacterial Filtration Efficiency (BFE) EN 14683:2019+AC:2019 Annex B			
A-Mask	X		
• Microbial cleanliness (bioburden) EN ISO 11737-1:2018			
A-Mask	X		
Breathability (Differential Pressure) EN 14683:2019+AC:2019 Annex C			
A-Mask	X		
Resistance against penetration by synthetic blood ISO 22609:2004			
A-Mask	X		

Remark: Test has been performed as per application request

Analytical Report Nr.

AR-21-YL-000131-01

Sample code Nr.

560-2020-00009768

Date

07/01/2021

COMPONENT LIST:

COMPONENT ID	COMPONENT NAME	MATERIAL DESCRIPTION	COLOR	REMARKS
CUST 01	A-Mask	Mask	Blue	---

Analytical Report Nr.

AR-21-YL-000131-01

Sample code Nr.

560-2020-00009768

Date

07/01/2021

MASKS TESTING	CAS No.	RESULTS	UNC.	LOQ	GUIDELINES
---------------	---------	---------	------	-----	------------

Analyses on:A-Mask
• Bacterial Filtration Efficiency (BFE)

Analysis Ending Date: 23/12/2020

EN 14683:2019+AC:2019 Annex B

Bacterial Filtration Efficiency (BFE)

99.81 %

- ≥ 98 %

✓ Pass

 Complete test report attached as Annex
 Test covered by ACCREDIA accreditation scope n° 1827 L

Breathability (Differential Pressure)

Analysis Ending Date: 18/12/2020

EN 14683:2019+AC:2019 Annex C

Differential pressure

 47.1 Pa/cm² (± 1.4) Pa/cm²

 - <60 Pa/cm²

✓ Pass

Complete test data reported at Annex.

Resistance against penetration by synthetic blood

Analysis Ending Date: 07/01/2021

ISO 22609:2004

Number of specimens tested

32

-

N° of specimens failed

3

-

N° of specimens passed

29

- ≥29

✓ Pass

At least 29 of 32 specimens must pass tested at 16KPa

Complete test data reported at Annex.

AQL information according to ISO 22609:2004:

A single sampling plan providing an AQL of 4,0 % requires 32 test specimens.

An AQL of 4.0 % is met for a single sampling plan when 29 or more specimens show pass results.

AQL= Acceptable Quality Limit.

Analytical Report Nr.

AR-21-YL-000131-01

Sample code Nr.

560-2020-00009768

Date

07/01/2021

MASKS TESTING	CAS No.	RESULTS	UNC.	LOQ	GUIDELINES
• Microbial cleanliness (bioburden)					Analysis Ending Date: 28/12/2020
EN ISO 11737-1:2018					
Bioburden		<30 cfu/g	-	≤ 30 cfu/g	✓ Pass

Complete test data reported at Annex.

Test covered by ACCREDIA accreditation scope n° 1827 L

Analytical Report Nr.

AR-21-YL-000131-01

Sample code Nr.

560-2020-00009768

Date

07/01/2021

Signed for and on behalf of Eurofins Textile Testing Spain:

Eurofins Textile Testing Spain, S.L.U.
C/ Castellón de la Plana (Alicante)
E-03070099

Report electronically validated by

Maria Jesus Martinez Puig

Chemical Lab manager

EXPLANATORY NOTE

- ◆ Test not covered by ENAC accreditation scope
- Test is subcontracted within Eurofins group and is accredited
- Test is subcontracted within Eurofins group and is not accredited
- Test is subcontracted outside Eurofins group and is accredited
- Test is subcontracted outside Eurofins group and is not accredited

N/A = Not Applicable

*Eurofins Textile Testing Spain S.L.U is not responsible of the information supplied by the customer and reported as section "Information provided by the customer".

Eurofins General Sales Terms and Conditions Applied.

Results obtained refer only to samples, products or material received in Laboratory, as described in section "Sample information" and tested in conditions shown in present report.

Test uncertainties not reported are at customer disposal, for those tests in which it is possible to evaluate the test uncertainty.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution provides a level of confidence of approximately 95%.

Reproduction of this document is only valid if it is done completely and under the written permission. Any printed version of this document will be considered as a copy.

If you happen to have any comments, please do it by sending email to textile_spain@eurofins.com and referring to this report number.

End Of Report

Eurofins Textile Testing Spain, S.L.U.

Calle Germán Bernácer, 4

03203 Elche

SPAIN

Phone+34 966 299 638**www.eurofins.com/tex**

ENAC is signatory of EA and ILAC Multilateral Agreement for testing
Activities not covered by ENAC accreditation are marked with ◆●○□■



Cosmetics &
Personal Care



LAB N° 1827 L


Page: 1 of 1

TEST REPORT	Refer to Analytical Report Number																																										
SPONSOR	Eurofins Textile & Footwear Testing Spain																																										
	C/Germán Bernácer 4																																										
	03203 Elche (Alicante)																																										
	SPAIN																																										
TEST METHOD	Microbial cleanliness (Bioburden) – EN 14683:2019/AC 2019 par. 5.2.5 + App D																																										
TEST ITEM - INFORMATION FROM THE SPONSOR																																											
PRODUCT NAME	560-2020-00009768 - masks																																										
MATRIX OF THE PRODUCT	Face Mask																																										
BATCH	EUAA70-00009731	CODE	Not provided																																								
EUROFINS COSMETICS & PERSONAL CARE ITALY IDENTIFICATION																																											
MATERIAL ITEM ALIQUOT	N720AA1118-1																																										
PARCEL REGISTRATION N.	IP-N7-2020350-AAF	RECEIVING DATE	15 Dec 2020																																								
ANALYSIS STARTING DATE	21 Dec 2020	ANALYSIS ENDING DATE	28 Dec 2020																																								
RESULTS	<table border="1"> <thead> <tr> <th>TOTAL BIOBURDEN</th> <th>SPECIFICATION</th> <th>RESULT</th> <th>UNIT</th> </tr> </thead> <tbody> <tr> <td rowspan="2">ALIQUOT 1</td> <td>/</td> <td>6.00</td> <td>CFU/sample</td> </tr> <tr> <td>≤ 30</td> <td>1.92</td> <td>CFU/g</td> </tr> <tr> <td rowspan="2">ALIQUOT 2</td> <td>/</td> <td><6.00</td> <td>CFU/sample</td> </tr> <tr> <td>≤ 30</td> <td><1.96</td> <td>CFU/g</td> </tr> <tr> <td rowspan="2">ALIQUOT 3</td> <td>/</td> <td><6.00</td> <td>CFU/sample</td> </tr> <tr> <td>≤ 30</td> <td><1.84</td> <td>CFU/g</td> </tr> <tr> <td rowspan="2">ALIQUOT 4</td> <td>/</td> <td><6.00</td> <td>CFU/sample</td> </tr> <tr> <td>≤ 30</td> <td><1.94</td> <td>CFU/g</td> </tr> <tr> <td rowspan="2">ALIQUOT 5</td> <td>/</td> <td><6.00</td> <td>CFU/sample</td> </tr> <tr> <td>≤ 30</td> <td><1.88</td> <td>CFU/g</td> </tr> </tbody> </table>				TOTAL BIOBURDEN	SPECIFICATION	RESULT	UNIT	ALIQUOT 1	/	6.00	CFU/sample	≤ 30	1.92	CFU/g	ALIQUOT 2	/	<6.00	CFU/sample	≤ 30	<1.96	CFU/g	ALIQUOT 3	/	<6.00	CFU/sample	≤ 30	<1.84	CFU/g	ALIQUOT 4	/	<6.00	CFU/sample	≤ 30	<1.94	CFU/g	ALIQUOT 5	/	<6.00	CFU/sample	≤ 30	<1.88	CFU/g
	TOTAL BIOBURDEN	SPECIFICATION	RESULT	UNIT																																							
	ALIQUOT 1	/	6.00	CFU/sample																																							
		≤ 30	1.92	CFU/g																																							
	ALIQUOT 2	/	<6.00	CFU/sample																																							
		≤ 30	<1.96	CFU/g																																							
	ALIQUOT 3	/	<6.00	CFU/sample																																							
		≤ 30	<1.84	CFU/g																																							
	ALIQUOT 4	/	<6.00	CFU/sample																																							
		≤ 30	<1.94	CFU/g																																							
	ALIQUOT 5	/	<6.00	CFU/sample																																							
		≤ 30	<1.88	CFU/g																																							

*This test report may not be reproduced in part unless expressly approved in writing by Eurofins Cosmetics & Personal Care Italy Srl.
The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor.
Information on the test item provided by the Sponsor are under Sponsor responsibility.*

Eurofins Cosmetics & Personal Care Italy Srl – via B.Buozzi 2, Vimodrone (Milano), Italy –
P.IVA / VAT Number: 05533561006
Tel: +39-022507151 – Fax: +39-0225071599 – E-mail: : InfoCosme@eurofins.com

Reviewed and electronically signed for Technical Supervisor Approval by
Martina Casini, Laboratory Manager
for Eurofins Cosmetic & Personal Care Italy Srl, on 28-Dec-2020 16:30:54 UTC+01:00

TEST REPORT	Refer to Analytical Report Number																				
SPONSOR	Eurofins Textile & Footwear Testing Spain																				
	C/Germán Bernácer 4																				
	03203 Elche (Alicante)																				
	SPAIN																				
TEST METHOD	Bacterial Filtration Efficiency (BFE) – EN 14683:2019+AC:2019 App B																				
TEST ITEM - INFORMATION FROM THE SPONSOR																					
PRODUCT NAME	560-2020-00009768																				
MATRIX OF THE PRODUCT	Face Mask																				
BATCH	EUAA70-00009731	CODE	Not provided																		
EUROFINS COSMETICS & PERSONAL CARE ITALY IDENTIFICATION																					
MATERIAL ITEM ALIQUOT	N720AA1119-1																				
PARCEL REGISTRATION N.	IP-N7-2020350-AAF	RECEIVING DATE	15 Dec 2020																		
ANALYSIS STARTING DATE	21 Dec 2020	ANALYSIS ENDING DATE	22 Dec 2020																		
EXPERIMENTAL CONDITIONS	Dimension of the test specimen: 175 mm x 95 mm Size of the area tested: 49 cm ² Flow rate during testing: 28,3 l/min Inner side of the mask to the aerosol challenge.																				
PHOTO OF THE TEST ITEM																					
RESULTS	<table border="1"> <thead> <tr> <th></th> <th>RESULT</th> <th>UNIT</th> </tr> </thead> <tbody> <tr> <td>ALIQUOT 1</td> <td>99,78</td> <td>%</td> </tr> <tr> <td>ALIQUOT 2</td> <td>99,82</td> <td>%</td> </tr> <tr> <td>ALIQUOT 3</td> <td>99,86</td> <td>%</td> </tr> <tr> <td>ALIQUOT 4</td> <td>99,78</td> <td>%</td> </tr> <tr> <td>ALIQUOT 5</td> <td>99,82</td> <td>%</td> </tr> </tbody> </table>				RESULT	UNIT	ALIQUOT 1	99,78	%	ALIQUOT 2	99,82	%	ALIQUOT 3	99,86	%	ALIQUOT 4	99,78	%	ALIQUOT 5	99,82	%
	RESULT	UNIT																			
ALIQUOT 1	99,78	%																			
ALIQUOT 2	99,82	%																			
ALIQUOT 3	99,86	%																			
ALIQUOT 4	99,78	%																			
ALIQUOT 5	99,82	%																			
DETAILED RESULTS	See Addendum N. 1 (1 page)																				

This test report may not be reproduced in part unless expressly approved in writing by Eurofins Cosmetics & Personal Care Italy Srl. The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor. Information on the test item provided by the Sponsor are under Sponsor responsibility.



Cosmetics &
Personal Care



LAB N° 1827 L

Page: 2 of 2

Addendum N.1

Started on: 21/12/2020

Batch: N720AA1119

Sample description: 560-2020-00009768

Lot Number: EUAA70-00009731

Negative Control Plate Counts

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	Mean
Negative Control (CFU)	0	0	0	0	0	0	0

*number of colonies adjusted with positive-hole correction table

Positive Controls Plate Counts

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	Total CFU
Size of particle (µm)	7,00	4,70	3,30	2,10	1,10	0,65	
Positive Control N.1 (CFU)	146	188	1078	670	406	279	2767
Positive Control N.2 (CFU)	153	312	1036	543	423	313	2780

*number of colonies adjusted with positive-hole correction table

Mean of the total plate counts of the two positive controls (CFU): 2774

Mean Particle Size (MPS)

	MPS
Positive Control N.1 (µm)	2,71
Positive Control N.2 (µm)	2,79
Mean (µm)	2,75

Test specimens Plate Counts

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	Total CFU
N720AA1119-1 - Aliquot 1	0	0	0	0	3	3	6
N720AA1119-1 - Aliquot 2	0	0	0	0	1	4	5
N720AA1119-1 - Aliquot 3	0	0	0	0	2	2	4
N720AA1119-1 - Aliquot 4	0	0	0	0	1	5	6
N720AA1119-1 - Aliquot 5	0	0	0	0	1	4	5

*number of colonies adjusted with positive-hole correction table

Test specimens Bacterial Filtration Efficiency (BFE)

	BFE (%)
N720AA1119-1 - Aliquot 1	99,78
N720AA1119-1 - Aliquot 2	99,82
N720AA1119-1 - Aliquot 3	99,86
N720AA1119-1 - Aliquot 4	99,78
N720AA1119-1 - Aliquot 5	99,82

Eurofins Cosmetics & Personal Care Italy Srl – via B.Buozzi 2, Vimodrone (Milano), Italy –

P.IVA / VAT Number: 05533561006

Tel: +39-022507151 – Fax: +39-0225071599 – E-mail: : InfoCosme@eurofins.com

METHOD FOR DETERMINATION OF BREATHABILITY (DIFFERENTIAL PRESSURE)

Test Method: EN 14683: 2019+AC: 2019 Annex C

Number of test specimens: 5

Number of test per specimen: 5

Sample area tested: Circular, diameter 2,5 cm


Tested area of the test sample: 4,9 cm²

Flow rate during testing: 8±0,6 l/min

General location of measurement areas: Representative of the overall surface.

Conditioning: T^a between 16,7°C and 26°C. RH between 82,8% and 88% during at least 4 h.

Airflow direction during testing: From the inner layer to the outer layer.



Results

Specimen	Units (Pa)					Mean value (Pa)	ΔP (Pa/cm ²)
	Position 1	Position 2	Position 3	Position 4	Position 5		
1	244	222	249	225	231	234	47,8
2	230	222	224	219	225	224	45,7
3	235	220	238	224	249	233	47,6
4	251	224	223	228	224	230	46,9
5	242	221	234	225	237	232	47,3
						Mean Value	47,1
						Uncertainty	± 1,4

Observation:

For thick and rigid masks the test method may not be suitable as a proper seal cannot be maintained in the sample holder.

DETERMINATION RESISTANCE AGAINST PENETRATION BY SYNTHETIC BLOOD

Test Method: ISO 22609:2004; Targeting-plate test method

Number of test specimens: 32

Sample size: Circular, diameter 5,58 cm

Sample area tested: 24,5 cm²

Pressure: 16 kPa (120,0 mm Hg)

Stream velocity of synthetic blood: 550±10 cm/s

Distance of the face mask target area surface from the tip of the cannula: 30,5 cm

Angle of the pneumatic valve with respect to the face mask target area: 90°

Technique used to enhance visual detection of synthetic blood: Hydrophilic cotton

Conditioning: At least 4 hours. T^a between 16,7°C and 26°C. RH between 82,8% and 88%

Environmental test conditions 17,5°C; 78,8% Hr

Pre-treatment: None

Specimen	Results	
	Pass	Fail
1	X	
2	X	
3	X	
4	X	
5	X	
6	X	
7	X	
8	X	
9	X	
10	X	
11	X	
12	X	
13	X	
14	X	
15	X	
16	X	
17	X	
18	X	
19		X
20	X	
21	X	
22	X	
23	X	
24	X	
25	X	
26		X
27	X	
28	X	
29		X
30	X	
31	X	
32	X	

Conclusion	PASS
-------------------	-------------

Operating requirements for surgical masks based on EN 14683: 2019+AC: 2019 standard

TEST	TYPE I	TYPE II	TYPE IIR
Bacterial filtration efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98
Differential pressure (Pa/cm ²)	< 40	< 40	< 60
Splash resistance pressure (kPa)	Not required	Not required	≥ 16
Microbial cleanliness (CFU/g)	≤ 30	≤ 30	≤ 30